



17TH ADVANCED BEAM DYNAMICS WORKSHOP ON

FUTURE LIGHT SOURCES

CVD Diamond Based XBPM for High-Power-Density X-ray Beams

D. Shu, ANL

APRIL 6-9, 1999

ARGONNE NATIONAL LABORATORY, ARGONNE, IL U.S.A.

CVD Synthetic diamond properties


- High thermal conductivity
- Low thermal expansion coefficient
- Good mechanical strength and stiffness
- High radiation stiffness
- X-ray transparency
- 5.5-eV band gap

CVD Synthetic diamond applications at the APS


- Filters and windows for undulator beamline
- Blade material for undulator beamline front end XBPM
- Photoelectron-emission-type x-ray-transmitting XBPM
- PSPCD as x-ray-transmitting XBPM
- PSPCD as x-ray-transmitting x-ray beam profiler

CVD Synthetic diamond based XBPM study for 4th generation LS


- Design optimization (photo-emission-type or PSPCD-type)
- Cooling structure for high peak power
- Integrated design with beamline front end components




1DF7U





VIEW A UNDULATOR FE FIRST XBPM




VIEW B UNDULATOR FE SECOND XBPM





APS PSPCD @ ESRF BL-6 TBPM






APS TBPM TEST AT ESRF




The Aluminum coating pattern
on the CVD diamond disk for TBPM

The APS CVD Diamond Transmitting XBPM (TBPM)
used in commissioning studies in 1-ID-A FOE




SAMPLE SET-UP CONFIGURATIONS FOR THE CVD-DIAMOND-BASED TRANSMITTING XBPM TEST AT THE APS AND ESRF

02099642




Output signal from a photoconductive-type TBPM test sample in configuration (d) as a function of the stage vertical motion. ID-6 undulator gap was 29-mm. ESRF storage ring current was 110-mA.


02119605




Output signal from a photoemission-type TBPM test sample in configuration (c) as a function of the stage vertical motion. ID-6 undulator gap was 28.2-mm. ESRF storage ring current was 121-mA

B2-20 and B2-30 x-ray beam-position-monitor/fixed mask combination for 4-ID






B2-98112103.BMP








B2-98112101.BMP



A typical profile of APS undulator white beam directly measured by a 16-pixel linear-array PSPCD.

APS 16 X 16 pixel Synthetic Diamond-Based
Position-Sensitive Photoconductive Detector (PSPCD) Prototype





Different images observed on the LED array with various undulator gap sizes and slits settings.

